

Discovery | Drug Development | Diagnostics

Product:	Anti-Gentamycin Antibody					
Synonyms:	Not available					
Immunogen:	Gentamycin-BTG					
	1 mg PAS9652	Host Species:	Ovis aries (Sheep)	Isotype:	IgG	
Pack sizes: Cat. No.	10 mg PAS9230	Host Breed:	Texel	Format:	Clear liquid	
		Clonality:	Polyclonal	pH:	7.4	
IgG Concentration: Lot dependent. Determined @ 280nm.		Buffer: 20mM Phosphate	Buffer: 20mM Phosphate, 150mM Sodium Chloride.			
Recommended Working Concentration*: 2.5 µg/mL 'The application notes include recommended starting dilutions; optimal dilutions/concentrations should be determined by the end user.		Preservative: 0.09% Sodium A	Preservative: 0.09% Sodium Azide.			
Assessment Method: Competitive ELISA.		Species Reactiv	Species Reactivity: N/A			
Method of Purification: Salt fractionation.		For long term sto Avoid repeated fr	Storage: Can be stored for up to 3 months at +2°C - +8°C. For long term storage, aliquot and store at ≤-20 °C. Avoid repeated freeze/thaw cycles. Product should be protected from light exposure.			
	d Applications: suitable for the development of	immunoassays or imr	munoaffinity purification	columns.		
Sensitivity: 5 ng/mL Gentamycin sulphate produces 90% inhibition in a competitive ELISA employing Gentamycin polyclonal antibody.		ELISA microtitre antibody exhibite Sulphate 27.1% a	Target Specificity*: ELISA microtitre plate based and expressed as % crossreactivity*. The antibody exhibited crossreactivity for the following compounds: Netilmicin Sulphate 27.1% and Sisomicin Sulphate 36.5%.			
Related Produ	cts:	* Cross reactivity profile may	vary with tracer used.			

Gentamycin-HRP, HRP9609, Kit size: 0.5 mL.

Notes & Precautions:

Antibody can be affinity purified on request. This product as supplied is intended for research applications only, not for use in therapeutic or diagnostic applications without the expressed written authorization of Randox BioReagents. A safety data sheet (SDS) can be supplied upon request. Vial should be centrifuged briefly before opening to ensure all material is removed from the vial cap.

28-Apr-22 Information correct at time of going to print

