

Discovery | Drug Development | Diagnostics

Product:	Anti-Norfentanyl Antibody						
Synonyms:	Not available						
Immunogen:			Norfe	Norfentanyl-BTG			
	1 mg	PAS10202	Host Species:	Ovis aries (Sheep)	Isotype:	IgG	
Pack sizes: Cat. No.	10 mg	PAS10203	Host Breed:	Texel	Format:	Clear liquid	
			Clonality:	Polyclonal	pH:	7.4	
IgG Concentration:			Buffer:	Buffer:			
Lot dependent. Determined @ 280nm.			20mM Phosphate	20mM Phosphate, 150mM Sodium Chloride.			
Recommended Working Concentration*:  2.5 µg/mL  'The application notes include recommended starting dilutions; optimal dilutions/concentrations should be determined by the end user.			Preservative: 0.09% Sodium A	Preservative: 0.09% Sodium Azide.			
Assessment Method:			Species Reactiv	Species Reactivity:			
Competitive ELISA.			N/A	N/A			
Method of Purification:			Storage:	Storage:			
Salt fractionation.			For long term sto Avoid repeated fr	Can be stored for up to 3 months at +2°C - +8°C. For long term storage, aliquot and store at ≤-20 °C. Avoid repeated freeze/thaw cycles. Product should be protected from light exposure.			
Recommended	l Applica	tions:	L				
The antibody is	suitable t	for the development of	immunoassays or imn	nunoaffinity purification	columns.		
Sensitivity:			Target Specificit	Target Specificity*:			
10 ng/mL Norfentanyl produces 94% inhibition in a competitive ELISA employing Norfentanyl polyclonal antibody.			antibody exhibite	ELISA microtitre plate based and expressed as % crossreactivity*. The antibody exhibited crossreactivity for the following compounds: Fentanyl 335.5% and Despropionylfentanyl <4.7%.			
Related Products:			* Cross reactivity profile may	* Cross reactivity profile may vary with tracer used.			

Norfentanyl-HRP, HRP9539, Kit size: 0.5 mL.

## Notes & Precautions:

Antibody can be affinity purified on request. This product as supplied is intended for research applications only, not for use in therapeutic or diagnostic applications without the expressed written authorization of Randox BioReagents. A safety data sheet (SDS) can be supplied upon request. Vial should be centrifuged briefly before opening to ensure all material is removed from the vial cap.

28-Apr-22 Information correct at time of going to print

